

Data evaluation memo

Date of evaluation: 7/20/2016

Person evaluating data: Erin S. McCoy, RPM

Site: Des Moines TCE Site

Laboratory: Pace Analytical Services, Inc. (Pace), Lenexa, KS 66219

SDG Number: 60221374

Number of samples: 6

Analytical method: SW846 1613B

National Functional Guidelines document used for evaluation: National Functional Guidelines for Chlorinated Dibenzo-p-Dioxins (CDDs) and Chlorinated Dibenzofurans (CDFs) Data Review, September 2011.

The above referenced SDG was evaluated to determine if samples were analyzed as specified by the Quality Assurance Project Plan (QAPP) for the project. The following were reviewed:

- ☐ Sample temperature at the time of lab receipt
- ☐ Hold times
- ☐ Blank contamination
- ☐ Laboratory spike and sample recoveries are within acceptable limits
- ☐ Laboratory narrative

Based on the review of the data, the data is acceptable without qualification. Further detail is provided below:

Sample temperature at the time of lab receipt – Samples were received at ~1.6-1.7 degrees Celsius, which is within the acceptable temperature between freezing and 6 degrees Celsius.

Hold times – Samples were collected on 6/10/2016. Samples were received on 6/14/16. Samples were extracted between 6/17-21/16. All samples were analyzed between 6/21-24/16. Samples were analyzed within the 40 day hold time window specified by Pace for the method used.

Blank contamination – Method, field and rinsate blanks were included in the SDG. No chemicals were detected in any of the blank samples.

Laboratory spike and sample recoveries are within acceptable limits – As shown on pages 18 through 20 of the SDG, all LCS recoveries were within acceptable limits.

Laboratory narrative – Only one detection was qualified as estimated “J”. This was because the detection equaled the reporting limit. No further qualification is necessary. Matrix spikes were not prepared for the SDG.